



MIRUS™ Controller

The assistant by your side

Technical Specifications

General specifications

HW options

MC-MC-ISO (Isoflurane)

MC-MC-SEVO (Sevoflurane)

MC-MC-DES (Desflurane)

Dimensions

<i>Dimensions (LxHxD)</i>	325 x 195 x 210 mm (12.8 x 7.6 x 8.2 in.)	<i>Weight</i>	9.0 kg (15.4 lb)
---------------------------	----------------------------------------------	---------------	------------------

Environmental conditions (Operation)

<i>Temp. range (ISO, SEVO)</i>	+10 - +40°C	<i>Temp. range (DES)</i>	+10 - +30°C
<i>Atm. pressure range</i>	700 - 1,060 hPa	<i>Equivalent altitude</i>	3,000 - 0 m (9,840 - 0 ft)
<i>Humidity range</i>	10 - 90% relative		

Environmental conditions (Storage)

<i>Temperature range</i>	-20 to +50°C	<i>Atm. pressure range</i>	500 - 1,060 hPa
<i>Equivalent altitude</i>	5,500 - 0 m (18,050 to 0 ft)	<i>Humidity range</i>	10 - 90% relative

Environmental conditions (Transport)

<i>Temperature range</i>	-20 to +70°C	<i>Atm. pressure range</i>	500 - 1,060 hPa
<i>Equivalent altitude</i>	5,500 - 0 m (18,050 to 0 ft)	<i>Humidity range</i>	10 - 90% relative

Device volume

<i>Noise level</i>	≤ 49 dB (A)	<i>Alarms at max. setting</i>	> 66 dB (A) (tone sequence)
<i>"Panic" alarms</i>	> 63 dB (A) (permanent)		

Classifications

<i>CE class according to 93/42/EEC</i>	IIb	<i>Protection class acc. To EN 60601-1</i>	I
<i>IP Code</i>	IP20	<i>Protection type acc. to EN 60601-1</i>	B

GMDN codes

<i>MC-MC-ISO</i>	36890	<i>MC-MC-DES</i>	36979
<i>MC-MC-SEVO</i>	36980	<i>UMDNS code</i>	10-144

Electrical supply

<i>Nominal voltage</i>	100 - 230 VAC ± 10%	<i>Frequency</i>	50 - 60 Hz ± 5%
<i>Power consumption</i>	< 75 VA	<i>Grounding</i>	Standard ground stud
<i>Internal backup</i>	Backup time: 15 min		

Electrical supply

<i>Internal reservoir</i>	Max. capacity = 270 ml Reserve capacity = 20 ml	<i>Status information</i>	On Screen
<i>Filling system</i>	Proprietary filling system, agent spec. acc. ISO 5360	<i>Filling capacity</i>	Maximum = 190 ml Max. overflow protect = 250ml

Display structure

<i>Screen</i>	Touchscreen, 5.7"	<i>Menu language</i>	English, German, French, Italian, Spanish, Polish
---------------	-------------------	----------------------	------------------------------------------------------

Units

<i>CO₂</i>	mmHg, Vol%, kPa	<i>Pressure (Paw, PEEP)</i>	mbar (for German), cmH ₂ O (for other languages)
<i>VA concentration</i>	Vol%		

Electromagnetic compatibility

Test parameters and limit values can be obtained from the manufacturer if required.

Controls and Ranges

Agent dosage

Isoflurane

<i>Range</i>	MAC 0.1 - 1.5 @ MVi 3.0 - 15.0 L/min where MAC 1 = 1.15 Vol% @ age = 40	<i>Increment</i>	0.1 MAC @ single taps on button 0.5 MAC @ pressing button > 1 sec
--------------	-------------------------------------------------------------------------------	------------------	----------------------------------------------------------------------------

Sevoflurane

<i>Range</i>	MAC 0.1 - 1.5 @ MVi 3.0 - 15.0 L/min where MAC 1 = 1.9 Vol% @ age = 40	<i>Increment</i>	0.1 MAC @ single taps on button 0.5 MAC @ pressing button > 1 sec
--------------	------------------------------------------------------------------------------	------------------	----------------------------------------------------------------------------

Desflurane

<i>Range</i>	MAC 0.1 - 1.5 @ MVi 3.0 - 15.0 L/min where MAC 1 = 6.7 Vol% @ age = 40	<i>Increment</i>	0.1 MAC @ single taps on button 0.5 MAC @ pressing button > 1 sec
--------------	------------------------------------------------------------------------------	------------------	----------------------------------------------------------------------------

Alarm settings

et CO₂ min / max

<i>Range</i>	15 - 150 mmHg / 2.0 - 19.7 Vol% / 2.0 - 20.0 kPa	<i>Increment</i>	1 mmHg / 0.5 Vol% / 0.1 kPa @ single taps on button 5 mmHg / 1.0 Vol% / 0.5 kPa @ pressing button > 1 sec
--------------	--------------------------------------------------------	------------------	--------------------------------------------------------------------------------------------------------------------

et ISO min / max

<i>Range</i>		<i>Increment</i>	0.1 Vol% @ single taps on button
<i>min</i>	0.0 - 5.0 Vol%		0.5 Vol% @ pressing button > 1 sec
<i>max</i>	0.5 - 5.0 Vol%		

et SEVO min / max

<i>Range</i>		<i>Increment</i>	0.1 Vol% @ single taps on button
<i>min</i>	0.0 - 8.0 Vol%		0.5 Vol% @ pressing button > 1 sec
<i>max</i>	0.5 - 8.0 Vol%		

et DES min / max

<i>Range</i>		<i>Increment</i>	0.1 Vol% @ single taps on button
<i>min</i>	0.0 - 12.0 Vol%		0.5 Vol% @ pressing button > 1 sec
<i>max</i>	0.5 - 12.0 Vol%		

Apnoe time

<i>Range</i>	15 - 60 sec	<i>Increment</i>	5 seconds
--------------	-------------	------------------	-----------

Alarm volume

<i>Range</i>	50 - 100%	<i>Increment</i>	10% @ single taps on button
--------------	-----------	------------------	-----------------------------

Patient data settings

Age

<i>Range</i>	10 - 115 years	<i>Increment</i>	1 year @ single taps on button 5 years @ pressing button > 1 sec
--------------	----------------	------------------	---------------------------------------------------------------------------

Size

<i>Range</i>	100 - 250 cm	<i>Increment</i>	5 cm @ single taps on button 10 cm @ pressing button > 1 sec
--------------	--------------	------------------	-----------------------------------------------------------------------

Weight (IBW)

<i>Range</i>	15 - 125 kg	<i>Increment</i>	1 kg @ single taps on button 5 kg @ pressing button > 1 sec
--------------	-------------	------------------	----------------------------------------------------------------------

Monitoring system

Safety monitoring

<i>Patient</i>	Breathing activity	<i>Alarms</i>	Apnoea
<i>Tidal volume</i>	Low Vt		Low Vti

Gas monitoring

ISO, SEVO, DES			
<i>Numerical data</i>	etVA	<i>Alarms</i>	Low etVA, High etVA
CO ₂			
<i>Numerical data</i>	etCO ₂	<i>Alarms</i>	Low etCO ₂ , High etCO ₂

Performance

Agent dosage accuracy

ISO, SEVO, DES			
<i>Maximum</i>	+ 15% of set MAC @ all settings	<i>Minimum</i>	- 15% of set MAC @ Vti = 500 ml, Rate = 15/min, I:E = 1:2, MAC = 1.0
<i>Wash-In speed</i>	Turtle	<i>Maximum Rate</i>	40/min

Respiratory monitoring accuracy

<i>Volumes</i>	± 20% @ Vt 200 ml - 2000 ml (BTPS corrected @rate < 40/min)	<i>Pressures</i>	± 4.0% or 2 cmH ₂ O whichever is greater @ -10 - 100 cmH ₂ O (@rate < 40/min)
<i>Rate</i>	± 1 breath per minute (@rate < 40/min)	<i>I:E</i>	± 25% (@rate > 40/min)

Gas monitoring accuracy

CO ₂	0 - 1 Vol% ± 0.1 Vol% 1 - 5 Vol% ± 0.2 Vol% 5 - 7 Vol% ± 0.3 Vol% 7 - 10 Vol% ± 0.5 Vol% > 10 Vol% unspecified	<i>Isoflurane</i>	0 - 1 Vol% ± 0.15 Vol% 1 - 5 Vol% ± 0.2 Vol% > 5 Vol% unspecified
<i>Sevoflurane</i>	0 - 1 Vol% ± 0.15 Vol% 1 - 5 Vol% ± 0.2 Vol% 5 - 8 Vol% ± 0.4 Vol% > 8 Vol% unspecified	<i>Desflurane</i>	0 - 1 Vol% ± 0.15 Vol% 1 - 5 Vol% ± 0.2 Vol% 5 - 10 Vol% ± 0.4 Vol% 10 - 15 Vol% ± 0.6 Vol% 15 - 18 Vol% ± 1.0 Vol% > 18 Vol% unspecified



Technologie Institut Medizin GmbH (TIM)
August-Thyssen-Str. 30
56070 Koblenz
Germany
Tel.: +49 261 899 689 00
Fax.: +49 261 899 689 09
www.tim-gmbh.de
info@tim-gmbh.de



Specifications are subjects to change without prior notice.

MIRUS™ Controller Technical specifications
©2022 TIM GmbH. All rights reserved.

ID No. 021-00208-C, 2022/04

MIRUS and TIM are trademarks of TIM GmbH.



Visit us
www.the-mirus.com