



Technologie Institut Medizin GmbH (TIM)  
August-Thyssen-Str. 30  
56070 Koblenz  
Deutschland

Subject: DOC TIM MIRUS

Date: 12. December 2019

**Declaration of Compatibility for the TIM MIRUS Controller / ARKON-PH's Interface with Philips IntelliBridge EC10 and EC40/80**

Dear Mr Kriesmer,

Philips has received the test and validation reports for your MIRUS Controller and ARKON PH Interface. Your document 063\_0038\_ARKON\_VnV-Result\_Eval\_A-00 revision A.00 dated 26-OCT-2019, is now filed in our Device History File (DHF). According to your test summary report, verification tests have passed successfully.

With successful testing and provision of your documentation you meet the necessary conditions for Philips to claim the TIM MIRUS Controller with ARKON PH interface a compatible and "supported device" for Philips Patient Monitoring Systems IntelliVue MP40-90 and MX400-850 using the Philips IntelliBridge EC10 Interface Module or IntelliBridge EC10 integral Interface Board with Open Interface driver (ED101) and Philips EC5 ID Module #101.

This DOC also includes compatibility with Philips IntelliBridge System Release C.0 and Patient Information Center iX using the EC40/80 Hub with Open Interface and EC5 ID Module #101.

Philips confirms that Technologie Institut Medizin GmbH (TIM) may now claim compatibility of the MIRUS Controller with ARKON Data Bridge (Philips Option) Software Version 1.00.00 with the above-mentioned Philips IntelliVue Patient Monitoring Systems.



**Philips Medizin Systeme Böblingen GmbH**

Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202  
Geschäftsführer: Dr. Dieter Haase (Sprecher), Dr. Peter Ziese, Matthijs de Groot  
Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, [www.philips.com/healthcare](http://www.philips.com/healthcare)

Philips is going to include your MIRUS Controller inhaled sedation system in the list of Open Interface supported medical devices that is available to the Philips sales channels and published on related Philips web pages.

Best regards,



Andreas Bindzus

Senior Product Manager, Medical Device Integration

Monitoring and Analytics